

Instructions for Use of Cell Residual Human IL-4 ELISA Detection Kit

The kit is intended for scientific research only and should not be used for diagnosis

Cat. No. HG- IL004

Introduction

This kit uses double-antibody sandwich enzyme-linked immunosorbent assay (ELISA) method. Add human IL-4 standard and test samples to the microtiter plate pre-coated with anti-human IL-4 antibody, then add diluted biotin-labeled human IL-4 detection antibody, finally add streptavidin-HRP to form the antibody + antigen + antibody-Biotin + SA-HRP complex, wash the plate and add TMB chromogenic solution for color development. TMB is converted from colorless to blue under the catalysis of HRP enzyme and finally to yellow under the action of stop solution. The shade of yellow is positively correlated with the amount of human IL-4 detected in the samples.

Assay range: 1.37-1000 pg/mL

Sensitivity: 0.28 pg/mL

Specification

96 T

Usage

It is applicable to the optimization of purification process of biological products, impurity control of intermediate process and release testing of final products.

Kit components

Components	Specification	Preparation
Standard	S, S1-S7, S0	Ready-to-use
Coated Plate	8 wells × 12 strips	Ready-to-use
Assay Buffer	12 mL × 1 vial	Ready-to-use
Wash Buffer (10x)	50 mL × 1 vial	Make a 10-fold dilution with ultrapure water.
Detection Antibody	6 mL × 1 vial	Ready-to-use
Streptavidin-HRP	12 mL × 1 vial	Ready-to-use
TMB Substrate	12 mL × 1 vial	Ready-to-use
Stop Solution	12 mL × 1 vial	Ready-to-use
Sealing Film	5 pieces	Ready-to-use
Instructions for Use	1 copy	Ready-to-use

Storage and shelf life


Sealed kits are valid for 12 months at 2-8°C.

Apparatus and materials to be prepared by the user:

- ◆ Plate reader
- ◆ Deionized water
- ◆ Thermostat plate shaker
- ◆ Unused filter paper
- ◆ Micro pipette and tips
- ◆ Vortex shaker

Pre-experiment preparation

1. All reagents and samples to be tested should be restored to room temperature. All reagents are newly prepared and ready-to-use.
2. Preparation of 1x washing solution: Equilibrate the concentrated washing solution to room temperature, without crystallization. After mixing well, according to the usage volume, dilute 10x washing solution by 10 times with an appropriate amount of ultrapure water at the ratio of 1:9, to obtain 1x washing solution.
3. Preparation of standards: Ready-to-use standards S1 to S7 and S0 can be used for serum and plasma samples. For other types of samples, prepare a standard curve with sample matrix (SPB, e.g., cell culture supernatant, tissue grind, cell lysate, etc. Urine samples are used to draw a standard curve using detection buffer. Standard dilution step: Human IL-4 high concentration standard S (10000 pg/mL), 30 μ L + 270 μ L SPB as a high standard (1000 pg/mL), add 200 μ L SPB to each dilution tube and dilute in a 1:2 gradient using high standard. Thoroughly mix each tube before performing the next transfer. SPB is used as zero standard (0 pg/mL).



Operation procedures

All reagent components and samples to be tested should be restored to room temperature before use. Duplicate well assay is recommended for all standards and samples to be tested.

1. Preparation of reagents: Prepare all reagents to be tested, diluted standards and samples to be tested in advance.
2. Microplate strip determination: Calculate the microtiter strips required for the samples to be tested and standards, remove the microtiter strips from the aluminum foil bag, place the remaining microtiter strips back into the aluminum foil bag and seal the mouth of the bag, and store it at low temperature.
3. Add 50 μ L detection buffer to each well.
4. Sample and test antibody incubation: Add the 50 μ L of standards and the sample to be tested to each well, and ensure that the spot sampling is completed within 15 min. Add 50 μ L of 1 x detection antibody to each well. Seal the plate with the sealing film and incubate in a 25°C thermostatic incubator at 500 rpm for 1 hour.
5. Plate washing: Discard the liquid in the wells, add 1 x washing solution (300 μ L/well) to wash the plate for 4 times, and pat dry the residual liquid in the microtiter plate.


6. Enzyme conjugate incubation: Add enzyme conjugate into microtiter plate with 100 μ L/well, seal the plate with sealing film, place it into a thermostatic incubator, incubate at 500 rpm for 30 minutes at 25 $^{\circ}$ C.
7. Plate washing: Same as Step 5.
8. Add 100 μ L of chromogenic substrate TMB to each well and incubate at room temperature for 15-20 minutes.
9. Termination: Add 100 μ L stop solution into each well, and gently shake the microtiter plate until the color development is uniform.
10. Readings: Read the absorbance value at 450 nm/630 nm within 20 minutes. Take 450 nm as detection wavelength and 630 nm as reference wavelength.

Results processing

1. OD processing of the standard curve (See the following example, which is only for example purpose. Please refer to the actual measurement for details):

Standard concentration (pg/mL)	OD value (1)	OD value (2)	Mean value
1000	3.8591	3.7755	3.8173
333.33	1.9249	1.9264	1.9257
111.11	0.7143	0.7335	0.7239
37.04	0.2498	0.2553	0.2526
12.35	0.0855	0.0907	0.0881
4.12	0.031	0.0371	0.0341
1.37	0.0163	0.0168	0.0166
0	0.0089	0.0085	0.0087

2. The standard curve is obtained by 4-parameter fitting with the theoretical standard concentrations and the corresponding OD values (as shown in the figure below).



Precautions

1. When the sample is tested for the first time, it is recommended to perform dilution with at least 3 consecutive dilution factors, so as to generate at least one diluted sample within the range of the standard curve.
2. The reagents should be stored as indicated on the label, and should be equilibrated to room temperature before use.
3. Before using the coated microtiter plates, please equilibrate to room temperature and then open the secondary packaging. The strip plates not used in the test should be immediately placed back into the package and sealed properly, and can be stored at 4°C for one month. Other unused reagents should be packaged or covered properly.
4. Please use disposable tips during experimental operation to avoid cross contamination.
5. Please check each individual reagent in the kit fully before use. To obtain accurate assay results, it is of special importance to mix well or shake well the reagents for dilution, loading, and reaction termination.
6. When washing residual Wash Buffer in the reaction wells, pat the plate dry adequately on clean tissue papers until watermark is no longer visible. Do not put the tissue paper into the well for liquid absorption.
7. The TMB Substrate is photosensitive, thus long-time exposure to illumination should be avoided; avoid contact with metal, otherwise, the assay results may be affected.
8. The kit is intended for single use. Please use within the shelf life.

Disclaimer

Under all circumstances, the liability of our company for this product is only limited to the value of the product itself.

